

Serum aldosterone was measured using AB Sciex 6500 liquid chromatography and tandem mass spectrometry (LC/MS/MS) at the University of Minnesota Pathology Laboratory. Limit of quantification was 4 ng/dL, at which point the intra-assay coefficient of variation was 13.9% (95% CI 7.8, 20.0).